



中国合格评定国家认可委员会 实验室认可证书

(注册号: CNAS L5066)

兹证明:

南通菲希尔测试仪器有限公司校准实验室

(法人: 南通菲希尔测试仪器有限公司)

上海市普陀区真北路 915 号 2208 室, 200333

符合 ISO/IEC 17025: 2017 《检测和校准实验室能力的通用要求》
(CNAS-CL01 《检测和校准实验室能力认可准则》) 的要求, 具备承担本
证书附件所列服务能力, 予以认可。

获认可的能力范围见标有相同认可注册号的证书附件, 证书附件是
本证书组成部分。


生效日期: 2019-04-18

截止日期: 2023-04-13



中国合格评定国家认可委员会授权人

中国合格评定国家认可委员会 (CNAS) 经国家认证认可监督管理委员会 (CNCA) 授权, 负责实施合格评定国家认可制度。
CNAS 是国际实验室认可合作组织 (ILAC) 和亚太实验室认可合作组织 (APLAC) 的互认协议成员。
本证书的有效性可登陆 www.cnas.org.cn 获认可的机构名录查询。



China National Accreditation Service for Conformity Assessment
LABORATORY ACCREDITATION CERTIFICATE
(Registration No. CNAS L5066)

Nantong Fischer Instrumentation Co., Ltd. Calibration Lab

(Legal Entity: Nantong Fischer Instrumentation Co., Ltd.)

Room 2208, No.915, Zhenbei Road, Putuo District, Shanghai, China

is accredited in accordance with ISO/IEC 17025: 2017 General Requirements for the Competence of Testing and Calibration Laboratories(CNAS-CL01 Accreditation Criteria for the Competence of Testing and Calibration Laboratories) for the competence to undertake the service described in the schedule attached to this certificate.

The scope of accreditation is detailed in the attached schedule bearing the same registration number as above. The schedule forms an integral part of this certificate.

Effective Date: 2019-04-18

Expiry Date: 2023-04-13

Signed on behalf of China National Accreditation Service for Conformity Assessment

China National Accreditation Service for Conformity Assessment(CNAS) is authorized by Certification and Accreditation Administration of the People's Republic of China (CNCA) to operate the national accreditation schemes for conformity assessment. CNAS is a signatory of the International Laboratory Accreditation Cooperation Mutual Recognition Arrangement (ILAC MRA) and the Asia Pacific Laboratory Accreditation Cooperation Mutual Recognition Arrangement (APLAC MRA). The validity of the certificate can be checked on CNAS website at <http://www.cnas.org.cn/english/findanaccreditedbody/index.shtml>

名称：南通菲希尔测试仪器有限公司校准实验室

地址：上海市普陀区真北路 915 号 2208 室

注册号：CNAS L5066

认可依据：ISO/IEC 17025:2017 以及 CNAS 特定认可要求

生效日期：2021 年 05 月 24 日 截止日期：2023 年 04 月 13 日

附件 5 认可的校准和测量能力范围

注：“测量仪器名称”栏仪器名称前标注*的项目可开展现场校准。

序号	测量仪器名称	被测量	校准规范	测量范围	扩展不确定度 ($k=2$)	说明	生效日期
1	*X 射线荧光镀层测厚仪	长度	X 射线荧光镀层测厚仪校准规范 JJF 1306	(0~0.3) μm	$U=0.022 \mu\text{m}$		2021-05-24
				(0.3~50) μm	$U_{\text{rel}}=8\%$		2021-05-24
2	*磁性、电涡流式覆层厚度测量仪	长度	磁性、电涡流式覆层厚度测量仪检定规程 JJG 818	(0~50) μm	$U=0.55 \mu\text{m}$	不测校准用厚度片	2021-05-24
				(50~1250) μm	$U_{\text{rel}}=1.1\%$		2021-05-24



No. CNAS L5066

在线扫码获取验证

Name: Nantong Fischer Instrumentation Co., Ltd. Calibration Lab

Address: Room 2208, No.915, Zhenbei Road, Putuo District, Shanghai, China

Registration No. CNAS L5066

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2021-05-24 Expiry Date: 2023-04-13

CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT
SCHEDULE OF ACCREDITATION CERTIFICATE

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
1	*X-Ray Fluorescence Coating Thickness Instrument	Length	C.S.for X-Ray Fluorescence Coating Thickness Instruments JJF 1306	(0~0.3) μm	U=0.022 μm		
				(0.3~50) μm	U _{rel} =8%		
2	*Magnetic And Eddy Current Measuring Instrument For Coating Thickness	Length	V.R. Of Magnetic And Eddy Current Measuring Instrument For Coating Thickness JJG 818	(0~50) μm	U=0.55 μm	Except for Coating Thickness Calibration Standard	
				(50~1250) μm	U _{rel} =1.1%		



No. CNAS L5066